


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVM-2-P

S.N.: 2078

Date: Nov. 5, 1984



- a) Output signal amplitude:  
0 TO +15 V.
- b) Pulse width:  
0 TO 2  $\mu$ SEC.
- c) Rise time:  
 $\leq 100$  PSEC.
- d) Fall time:  
 $\leq 135$  PSEC.
- e) PRF:  
0 TO 25  $\mu$ MHZ.
- f) Jitter, stability:  
OK
- g) Prime power:

10 VOLTS/DIV  
1.0  $\mu$ SEC/DIV  
5 MHz

+24  $\pm$  1.0 VOLTS  
530 mA (AMPA)  
*[Signature]*