

PULSE GENERATOR  
PERFORMANCE CHECK

Model:


AVMM-1-C-DP

S.N.:

2094

Date:

NOV 28 84



- a) Output signal amplitude:  
0 TO +50 VOLTS.
- b) Pulse width:  
0 TO 1 USEC
- c) Rise time:  
≤ 100 PSEC
- d) Fall time:  
≤ 135 PSEC
- e) PRF:  
0 TO 25 MHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
110/220V,  
50-60 HZ.

5 VOLTS/DIV  
1.0 USEC/DIV  
5.0 MHZ

*[Signature]*