


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-8V: 1V2-C-PW-M*

S.N.: *2095*

Date: *JAN 5 88*




- a) Output signal amplitude:
- b) Pulse width:
- c) Rise time:
- d) Fall time:
- e) PRF:
- f) Jitter, stability:
- g) Prime power:

10 VOLTS/DIV
1.0 nSEC/DIV
100 kHz

PULSE GENERATOR

PERFORMANCE CHECK

Model: ANP-AN-14V2-C-PN-M
S.N.: 2095
Date: DEC 4 84



- d) Output signal amplitude:
0 TO ± 20 VOLTS
- e) Pulse width:
0 TO 2 μ SEC.
- f) Rise time:
 ≤ 100 PSEC
- g) Fall time:
 ≤ 200 PSEC
- h) PRF:
0 TO 1000/KHZ
- i) Jitter, stability:
OK
- j) Prime power:

10 VOLTS/DIV
1.0 μ SEC/DIV
100 KHZ

110/220 V
50-60 HZ
JK