


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-L1

S.N.: 2117

Date: DEC 14 '84



- a) Output signal amplitude:  
71 ± 40 VOLTS.
- b) Pulse width:  
1 TO 100 NS/2
- c) Rise time:  
≤ 300 PS/2
- d) Fall time:  
< 2 NS/2
- e) PRF:  
0 TO 50 KHZ  
SEE FIG 1 d2
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 ± 1.0 VOLTS  
75 mA (MAX)

1.0 NS/2/DIV

3.2 VOLTS/DIV

50 KHZ

*[Signature]*