


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-L1

S.N.: 2120

Date: DEC 14 84



- a) Output signal amplitude:
 $\geq +40V$
- b) Pulse width:
 1 to 100 NSER
- c) Rise time:
 ≤ 300 PSEER
- d) Fall time:
 ≤ 2 NSER
- e) PRF:
 0 TO 50 KHz
- f) Jitter, stability:
~~500~~ $1+2$
- g) Prime power:
OK

$+15 \pm 1.0$ VDC
 75 mA (MAX)

1.0 NSER/DIV
 32 VONP/DIV
 50 KHz.

[Signature]