


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A.1-PW-N*

S.N.: *2121*

Date: *JAN 7 85*



- a) Output signal amplitude
0 TO -200V
- b) Pulse width:
50 NS62 TO 100 US62
- c) Rise time:
~ 10 NS62
- d) Fall time:
~ 10 NS62
- e) PRF:
0 TO 100 KHZ
- f) Jitter stability
0.5% MAX DUTY CYCLE
- g) Drainage power
OK

+ 24 ± 1.0 VOLT
450 mA (MAX)