


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-HVI-P

S.N.: 2129

Date: SEPT 19, 1984



- d) Output signal amplitude:
0 to +100 V
- e) Pulse width:
≤ 1 μsec.
- f) Rise time:
NA
- g) Fall time:
NA
- h) PRF:
0 to 100 kHz
- i) Jitter, stability:
OK
- j) Prime power:

100 VOLTS/DIV
1.0 μSEC/DIV
50 kHz

+15 ± 1.0 V
150 mA (MAX)
[Signature]