

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AMP-AN-15-C-EA-EW-PN-M

S.N.:

2133

Date:

NOV 30 1984



a) Output signal amplitude:

0 TO  $\pm 5$  VOLTS

b) Pulse width:

0 TO 4 NSER

c) Rise time:

$\leq 70$  PSEZ (P)

d) Fall time:

( $\leq 100$  PSEZ FOR -N)

$\leq 135$  PSEZ

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

110/220V,

50-60 Hz.

1.0 NSER/DIV

5 VOLTS/DIV

500 KHz.