


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVM-2-C-PN-D-M

S.N.: 2139

Date: NOV 5, 1984



- a) Output signal amplitude:
0 TO $\pm 15V$
-) Pulse width:
0 TO 2 μ SEC
-) Rise time:
 ≤ 100 PSEC
-) Fall time:
 ≤ 135 PSEC
-) PRF:
0 TO 25 MHz
-) Jitter, stability:
OK
-) Prime power:

10 VOLTS/DIV
1.0 μ SEC/DIV
5 MHz

110/220 V
50-60 Hz

CAUTION UNIT
SHIPPED WITH
VOLTAGE SELECTOR
CARD IN THE
120 V POSITION