


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-1-C-PN-05-M-CA-CU

S.N.: 2166

Date: FEB 4 1955



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
0 TO 4 NSER.
- c) Rise time:
 ≤ 100 PSEZ
- d) Fall time:
 ≤ 135 PSEZ
- e) PRF:
0 TO 1000 Hz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS / DIV
1.0 NSER / DIV
100 KHz.

120 / 240 V
50 - 60 Hz

[Handwritten signature]