


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVMN-1-P .

S.N.: 2170

Date: DEC. 13, 1984



d) Output signal amplitude:

0 TO +5V ...

e) Pulse width:

0 TO 6 NSEC.

f) Rise time:

≤ 100 PSEC

g) Fall time:

≤ 200 PSEC

h) PRF:

0 TO 50 MHz

i) Jitter, stability:

OK

j) Prime power:

+24 \pm 1.0 V

500 mA (MAX)

5 VOLTS/DIV
1.0 NSEC/DIV
5.0 MHz