


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-15-C-P*

S.N.: *2179*

Date: *FEB 18 85*



- a) Output signal amplitude:
0 TO + 5V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 - 60 kHz

1.0 NSEC/DIV
5.0 VOLTS/DIV
100 KHz

[Signature]