


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-5-C-P-M*
S.N.: *2191*
Date: *JAN 14 85*



- a) Output signal amplitude:
0 to +28 Amps
- b) Pulse width:
5 to 100 nsec.
- c) Rise time:
≤ 3 nsec
- d) Fall time:
≤ 4 nsec
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz.

= 30 Amps/div.
10 nsec/div.
R_L = 3 Ω.
4 kHz.

[Signature]