


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-M-142-PS-PN-M*

S.N.: *2212*

Date: *FEB 5 1985*



- a) Output signal amplitude:  
*0 TO ±20 VOLTS*
- b) Pulse width:  
*0 TO 2 NSOR*
- c) Rise time:  
*≤ 100 PSEZ*
- d) Fall time:  
*≤ 200 PSEZ*
- e) PRF:  
*0 TO 1 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*

*10 VOLTS/DIV*  
*1.0 NS/DIV*  
*100 KHz*

*[Signature]*