


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-APS-PN-M*

S.N.: *2224*

Date: *FEB 11 85*



- a) Output signal amplitude:
0 TO ± 15 VOLTS
- b) Pulse width:
4 TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
*0 TO 1 MHz,
SEE PAGES 1 & 2*
- f) Jitter, stability:
O.K.
- g) Prime power:

*120/240V
50-60Hz*

*10 VOLTS/DIV
1.0 NSEC/DIV
1.00 KHz*