

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AM-2-C-N-M-D

S.N.:

2243

Date:

DEC - 19 84



a) Output signal amplitude:

0 TO -15V...

b) Pulse width:

0 TO 2 NSER

c) Rise time:

≤ 100 PSEZ

d) Fall time:

≤ 135 PSEZ

e) PRF:

0 TO 25000 Hz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50 - 60 Hz

10 VOLTS/DIV

1.0 NSER/DIV

5.0 mA

*[Signature]*