


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AMP-V-1

S.N.: 2261

Date: JAN 14 85



- a) Output signal amplitude:  
7 ± 1.0 VOLTS
- b) Pulse width:  
0.2 TO 4.0 NSEC.
- c) Rise time:  
≤ 100 PSEC
- d) Fall time:  
≤ 135 PSEC
- e) PRF:  
0 TO 1000 Hz
- f) Jitter, stability:  
OK
- g) Prime power:

10 VOLTS/DIV  
1.0 NSEC/DIV  
1.00 KHz

+15 ± 1.0 VOLTS  
250 mA (MAX)