

PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-U-HV1-C-PN-M

S.N.: 2264

Date: MAR 21 85

a) Output signal amplitude:

0 TO ± 50 VOLTS

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

≤ 250 PSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 20 KHZ


f) Jitter, stability:

OK

g) Prime power:

120/240V

5.0 - 60 HZ



32 VOLTS/DIV

2.0 NSEC/DIV

20 KHZ