


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-HV3-C-05-PN

S.N.: 2284

Date: APRIL 24 85



- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
0 TO 100 NSER.
- c) Rise time:
 < 200 PSEER.
- d) Fall time:
 ≤ 2 NSER
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60Hz

100 VOLTS/DIV
5 NSER/DIV
5 KHz.