


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVPMMN-C-PN-M*

S.N.: *2321*

Date: *MAY 6 1985*



*5 VOLTS/DIV  
1.0 NSEC/DIV  
10 MHz*



*5 VOLTS/DIV  
5 NSEC/DIV  
50 MHz*

- a) Output signal amplitude:  
*0 TO ± 2 VOLTS*
- b) Pulse width:  
*0 TO 6 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 VOLTS  
50 - 60 MHz*