


PULSE GENERATOR
PERFORMANCE CHECK

Model: PNP-17V-15-C-PN-05-07A-02

S.N.: 2344

Date: MAY 28 85



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0 TO 4 NS
- c) Rise time:
P: ≤ 60 PSEC
M: ≤ 100 PSEC
- d) Fall time:
 ≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50 - 60 Hz

5 VOLTS/DIV
1.0 NS/DIV
5.00 kHz