


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-P

S.N.: 2366

Date: MAR 20 85



- a) Output signal amplitude:  
0 to +30 VOLTS
- b) Pulse width:  
10 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 to 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 ± 1.0 V  
250 mA (MAX)

16 VOLTS/DIV  
10 NSEC/DIV  
500 KHZ