


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-2-C-P-M

S.N.: 2379

Date: APRIL 16 85



- a) Output signal amplitude:
0 TO + 10 AMPS
- b) Pulse width:
0 TO 15 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 7.5 AMPS/DIV
5.0 NSEC/DIV
20 KHz

120/240 VOLTS
50-60 Hz

AK