


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*

S.N.: *2387*

Date: *ANNY 7 85*



- a) Output signal amplitude:
25 TO 100 AmPS
- b) Pulse width:
~25 μ SEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:

*~ 60 AmPS DIU
10 NSEC DIU
1 kHz*

+ 15 \pm 1.0 VOLT

200 mA (MAX)