


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-P
S.N.: 2390
Date: MAY 2 85



- a) Output signal amplitude:
0 TO +350 VOLTS
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 1.5 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

320 VOLTS/DIV
10 NSEC/DIV
2 KHZ

120/240 V
50-60 HZ
[Signature]