


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ACMAN 3A-C-N*

S.N.: *2391*

Date: *APR 2 85*



- a) Output signal amplitude:
0 TO 5 V DMB
- b) Pulse width:
5 TO 40 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 1 NSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 V DMB/DIV
10 NSEC/DIV
R_L = 50 Ω

120/240 V
50-60 Hz