

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVO-5B-C-N

S.N.:

2404

Date:

MAY 21 85



a) Output signal amplitude:

0 TO - 48 AMPS

b) Pulse width:

25 TO 500 NSEC

c) Rise time:

≤ 10 NSEC

d) Fall time:

≤ 10 NSEC

e) PRF:

0 TO 1 kHz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V

50 - 60 Hz

10 VOLTS/DIV

100 NSEC/DIV

$R_L = 0.7$ AMPS

1 kHz