


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUR-A-1-PW-C-P*

S.N.: *2412*

Date: *MAY 30 85*



- a) Output signal amplitude:
0 TO +200V TO 50- Ω
- b) Pulse width:
50 NSEC TO 10 μ SEC
- c) Rise time:
*(0.5% MAX DUTY CYCLE)
 \leq 10 NSEC*
- d) Fall time:
 \leq 10 μ SEC
- e) PRF: *0 TO 100 KHZ*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 HZ

50 VOLTS/DIV
100 NSEC/DIV
10 KHZ
R_L = 50 Ω

[Signature]