


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-1-P*

S.N.: *2416*

Date: *APRIL 17 85*



- a) Output signal amplitude:
0 to +5V
- b) Pulse width:
0 to 6 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 200 psec
- e) PRF:
0 to 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:.

+24 ± 1.0 V

500 mA (MAX)

5 VOLT/DIV
1.0 nsec/DIV
4.0 MHz