


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMN-1-N*

S.N.: *2417*

Date: *APRIL 17 85*



- a) Output signal amplitude:  
*0 to -5 VOLT*
- b) Pulse width:  
*0 to 6 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC.*
- e) PRF:  
*0 to 50 MHz*
- f) Jitter, stability:  
*OK.*
- g) Prime power:  
*OK.*

*5 VOLT/DIV  
1.0 NSEC/DIV  
40 mA*

*+24 ± 1.0 V  
500 mA (MAX)*  
*[Signature]*