


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *ANMAN-3A-N*

S.N.: *2421*

Date: *Jul 7 31 85*



Ⓐ *2 nsec/div*  
*5 volts/div*  
*5 MHz*



Ⓑ *5 nsec/div*  
*5 volts/div*  
*50 MHz*

a) Output signal amplitude:

*0 to -5V*

b) Pulse width:

*5 to 40 nsec*

c) Rise time:

*≤ 1 nsec*

d) Fall time:

*≤ 1 nsec*

e) PRF:

*0 to 50 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 ± 1.0 volt*

*200 mA (max)*