


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-D-PN*

S.N.: *2429*

Date: *MAY 14 '85*



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz

5 VOLTS/DIV
1.0 NSEC/DIV
5.0 MHz

[Signature]