


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-C-17*

S.N.: *2430*

Date: *APRIL 30 85*



- a) Output signal amplitude:
+125 TO +100 AMPS
- b) Pulse width:
≈ 25 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.5 KHZ.
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz.

*≈ 60 AMPS/DIV
10 NSEC/DIV
1.3 KHZ.*