


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3-P*

S.N.: *2505*

Date: *Jul 5 85*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
2 TO 15 NSEC.
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 600 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*+24 TO
+28 V
400 mA (MAX)*

*10 VOLTS/DIV
2 NSEC/DIV
5.0 MHz*