


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-M-15-C-N-PN-M*
S.N.: *2507*
Date: *AUG 8 85*



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK.
- g) Prime power:

5 VOLTS/DIV
1.0 NSEC/DIV
500 KHz

120/240V
50-60 Hz.