


PULSE GENERATOR
PERFORMANCE CHECK

Model: PVAN-1-P

S.N.: 2516

Date: Aug 5 1985



- a) Output signal amplitude: 0 TO +5 VOLTS
- b) Pulse width: 0 TO 6 NSEC.
- c) Rise time: ≤ 100 PSEC
- d) Fall time: ≤ 200 PSEC
- e) PRF: 0 TO 50 MHz
- f) Jitter, stability: OK
- g) Prime power:

5 VOLTS/DIV
5 NSEC/DIV
(≈ 40 MHz)

+24 \pm 1.0 VOLTS
450 mA (MAX)