


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-TN-1*

S.N.: *2525*

Date: *JULY 22 85*



*≈ 100 VOLTS/DIV
10 NSEC/DIV*

- a) Output signal amplitude:
-230 VOLTS + +30 VDC
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 2 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*120 / 240 V
50 - 60 Hz*

- h) PROPAGATION DELAY
≤ 100 NSEC