

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMV-1-P*

S.N.: *2529*

Date: *JUNE 19 85*

CRP
mm

- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 ± 1.0 V
500 mA (MAX)

5 VOLTS / DIV
5 NSEC / DIV
≈ 40 MHz

[Signature]