


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ADMMN-1-N*

S.N.: *2531*

Date: *JUNE 18 1985*



- d) Output signal amplitude:  
*0 TO -5V*
- e) Pulse width:  
*0.2 TO 6 NSEC*
- f) Rise time:  
*≤ 100 PSEC*
- g) Fall time:  
*≤ 200 PSEC*
- h) PRF:  
*0 TO 50 MHz*
- i) Jitter, stability:  
*OK*
- j) Prime power:

*5.0 VOLTS/DIV*  
*5.0 NSEC/DIV*  
*≈ 40 MHz*

*+24 ± 1.0 VOLT*  
*500 mA (max)*