


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-24-C-N*  
S.N.: *2540*  
Date: *JUNE 7 '85*



- a) Output signal amplitude:  
*0 TO - 10 AMPS*
- b) Pulse width:  
*0 TO 50 NSEC*
- c) Rise time:  
*≤ 1 NSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 20 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 7 AMPS/DIV*  
*5.0 NSEC/DIV*  
*10 KHZ*  
*R<sub>L</sub> = 5.1Ω*

*120/240V*  
*50-60 Hz*