


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-A-C-D-7N*

S.N.: *2545*

Date: *SEPT 15 85*



- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*4 TO 100 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 1 MHz (PW = 4 NSEC)*  
*0 TO 20 KHz (PW = 100 NSEC)*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV*  
*2.0 NSEC/DIV*  
*50 KHz.*

*120/240 V*  
*50-60 Hz.*