


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AN-1403-CAEW-P*

S.N.: *2550*

Date: *SEPT 12 85*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0 TO 2 NSEC.
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 250 PSEC.
- e) PRF:
0 TO 1000 Hz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 32 VOLTS/DIV
1.0 NSEC/DIV
100 KHz

+15 ± 1.0 VOLT
250 mA (MAX)
[Signature]