


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMH-S-1-P-A*

S.N.: *2556*

Date: *SEPT 20 85*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 135 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
1.0 NSEC/DIV
500 KHZ*

*+15 ± 1.0 VOLT.
250 mA (MAX)
*[Signature]**