


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A1-P4-PS-P*

S.N.: *2570*

Date: *Aug 16 '85*



-) Output signal amplitude:
0 TO +200 VOLTS
-) Pulse width:
50 NSEC TO 1.0 USEC
-) Rise time:
≤ 10 NSEC
-) Fall time:
≤ 10 NSEC
- e) PRF:
*0 TO 100 KHZ,
(0.57 MAX DUTY CYCLE)*
- f) Jitter, stability:
OK
- g) Prime power:

*50 VOLTS/DIV
100 NSEC/DIV*

*120/240 V
50-60 HZ*