


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-C-P*

S.N.: *2593*

Date: *SEPT 15 85*



- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0 TO 6 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS / DIV.  
1.0 NSEC / DIV  
5 MHz*

*120 / 240 V  
50 - 60 Hz*  
*[Signature]*