


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-A-1-PW-C-P*

S.N.: *2605*

Date: *SEPT 25 85*



- a) Output signal amplitude:  
*0 TO +200 VOLTS*
- b) Pulse width:  
*50 NSEC TO 1.0 USEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*0 TO 100 KHz.*
- f) Jitter, stability:  
*0.5% MAX DUTY CYCLE*
- g) Prime power:  
*OK*

*50 VOLTS/DIV*  
*200 NSEC/DIV*  
*10 KHz.*  
*R<sub>L</sub> = 50 Ω*

*120/240 V*  
*50-60 Hz.*