


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-14V3-C-DN*

S.N.: *2606*

Date: *SEPT 11 85*



- a) Output signal amplitude:
0 TO ±40 VOLTS
- b) Pulse width:
0 TO 2 NSEC.
- c) Rise time:
< 150 PSEC
- d) Fall time:
≤ 250 PSEC.
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*≈ 32 VOLTS/DIV
1.0 NSEC/DIV
100 KHz.*

*120/240V
50-60 Hz.*

[Signature]