


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PAH-N.*

S.N.: *2676*

Date: *OCT 30 85*



*16 VOLTS/DIV
1.0 NSEC/DIV
500 KHZ*

- a) Output signal amplitude:
0 TO -30 VOLTS.
- b) Pulse width:
≤ 1 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 ± 1.0 VOLT.
150 mA (MAX)*