


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMH-S P*

S.N.: *2681*

Date: *AUG 13 85*



- a) Output signal amplitude:
0 TO +15 V.
- b) Pulse width:
≤ 500 pSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

1.0 VOLT/DIV
1.0 NSEC/DIV
1.0 MHz

+15 ± 1.0 VOLT
150 mA (MAX)

[Signature]