


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-1-C-P*

S.N.: *2699*

Date: *DEC 5 85*



*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*100 KHZ*

- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*[Signature]*